



Research Article

NEW SCREENING TECHNIQUE FOR HEART FAILURE

Abhijit Ray

Heart Failure Specialist, Fortis Escorts Heart Institute, New Delhi

ARTICLE INFO

Article History:

Received 06th May, 2023

Received in revised form 14th June, 2023

Accepted 23rd July, 2023

Published online 28th August, 2023

ABSTRACT

Heart Failure is one of the most serious cardiovascular diseases across the world with a prevalence of close to 64 million patients across the globe. While diagnosis and prognosis of Heart Failure predominantly depends on LVEF, it has quite a few backdrops. A novel technology has emerged (patented by USPTO) wherein LVEF can be calculated from ECG graphs. This is sure to have a huge impact on healthcare at a global scale once the technology will be adapted into the modern-day ECG, Holter and Cardiac Monitoring machines.

Key words:

Heart Failure

Copyright© The author(s) 2023. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

INTRODUCTION

Heart failure has been singled out as an epidemic and is a staggering clinical and public health problem, associated with significant mortality, morbidity, and healthcare expenditures, particularly among those aged 65 and older.

Heart failure (HF) has been defined as a clinical syndrome with symptoms and or signs caused by a structural and/or functional cardiac abnormality and corroborated by elevated natriuretic peptide levels and/or objective evidence of pulmonary or systemic congestion. A new and revised classification of HF according to left ventricular ejection fraction (LVEF) includes HF with reduced EF (HFrEF): HF with an LVEF of $\leq 40\%$; HF with mildly reduced EF (HFmrEF): HF with an LVEF of 41% to 49%; HF with preserved EF (HFpEF): HF with an LVEF of $\geq 50\%$; and HF with improved EF (HFimpEF) : HF with a baseline LVEF of $\leq 40\%$, a ≥ 10 -point increase from baseline LVEF, and a second measurement of LVEF of $>40\%$ ¹.

Thus, LVEF is clearly the single-most important parameter for diagnosis and prognosis of Heart Failure cases. According to the American Heart Association, LVEF is measured (expressed as a percentage) of how much blood the left ventricle pumps out with each contraction. It is the most common method for assessment of systolic function. In many cardiovascular emergencies including Acute Myocardial Infarction, Complete Heart Block and Acute Congestive Cardiac Failure, there is often an urgency to measure LVEF since treatment depends on it.

At present, the standard method of assessing LVEF is by conducting an Echocardiography, using the modified Quinones Equation. Hence, LVEF can only be assessed as snapshots rather than as a continuous parameter. LVEF can be

determined using several invasive and non-invasive imaging modalities, either subjectively by visual estimation or objectively by quantitative methods i.e. echocardiography, magnetic resonance imaging (MRI), computed tomography (CT), gated equilibrium radionuclide angiography (commonly referred to as multiple-gated acquisition [MUGA] scan) and gated myocardial perfusion imaging with either single-photon emission computed tomography (SPECT) or positron emission tomography (PET). LVEF can also be measured non-invasively using the firstpass radionuclide technique, but this technique is rarely performed in the current era. However, none of these have the advantage of continuous monitoring. Apart from chronic factors like Hypertension, Diabetes, smoking etc, multiple acute factors also affect LVEF including

- Sympathetic stimulation
- Blood volume
- Respiration
- Environmental temperature²

Due to these, LVEF is a dynamic entity whose continuous monitoring might be extremely beneficial for treatment instead of snapshot values of LVEF (as is available from current LVEF measurement techniques).

Other problems faced by the current methods of LVEF measurement (specially Echocardiography) are:

1. LVEF estimation from Echocardiography is subject to observer bias
2. It requires specialised equipment (Echo machine) and trained personnel to obtain the results
3. It is costly, specially in cases where multiple readings are required

In view of these problems, an algorithm has been invented and patented from the United States Trademark and Patent Office

*Corresponding author: Abhijit Ray

Heart Failure Specialist, Fortis Escorts Heart Institute, New Delhi

(USPTO) which can calculate LVEF from an ECG graph. This has a number of advantages.

1. Since it is a computer-based reading, no human bias will be involved
2. It can give 24×7 readings of LVEF
3. Will not require any additional equipment or training to obtain the results
4. Extremely cheap when compared to the modern methods.

Adaptation of this technology means LVEF can now be displayed on ECG strips, 24 hour LVEF reports on Holter reports, and 24×7 monitoring of LVEF in ICU/CCU/OT setups where patients are already on ECG chest leads.

Possible Benefits

Approximately, HF affects more than 64 million people worldwide, 1 Million Heart Failure cases are diagnosed annually in India alone³. As of 2020, the Global market for Electrocardiography was expected to close at \$8.3 Billion with a CAGR of 8.3%⁴. According to a survey in 2019, the Indian ECG market alone was estimated to be around Rs.164.25 Crore with high-end machines accounting for over Rs.96 Crore alone.

How fast this innovation will improve the cardiovascular market in India is yet to be seen, however many manufacturers are surely in line to adapt this new algorithm into their machines.

References

1. Bozkurt B, Coats AJ, Tsutsui H, Abdelhamid M, Adamopoulos S, Albert N, Anker SD, Atherton J, Böhm M, Butler J, Drazner MH, Felker GM, Filippatos G, Fonarow GC, Fuuzat M, Gomez-Mesa JE, Heidenreich P, Imamura T, Januzzi J, Jankowska EA, Khazanie P, Kinugawa K, Lam CSP, Matsue Y, Metra M, Ohtani T, Francesco Piepoli M, Ponikowski P, Rosano GMC, Sakata Y, Seferović P, Starling RC, Teerlink JR, Vardeny O, Yamamoto K, Yancy C, Zhang J, Zieroth S. Universal Definition and Classification of Heart Failure: A Report of the Heart Failure Society of America, Heart Failure Association of the European Society of Cardiology, Japanese Heart Failure Society and Writing Committee of the Universal Definition of Heart Failure. *J Card Fail.* 2021 Mar 1;S1071-9164(21)00050-6. doi: 10.1016/j.cardfail.2021.01.022. Epub ahead of print. PMID: 33663906.
2. Ray, A; Chandana, T. PS-C20-5: WARM IV INFUSION IN HYPERTENSIVE URGENCY. *Journal of Hypertension* 41(Suppl 1):p e394, January 2023. | DOI: 10.1097/01.hjh.0000916856.58986.27
3. Yan, T., Zhu, S., Yin, X., Xie, C., Xue, J., Zhu, M., Weng, F., Zhu, S., Xiang, B., Zhou, X., Liu, G., Ming, Y., Zhu, K., Wang, C., & Guo, C. (2023). Burden, trends, and inequalities of heart failure globally, 1990 to 2019: A secondary analysis based on the global burden of disease 2019 study. *Journal of the American Heart Association*, 12(6). <https://doi.org/10.1161/jaha.122.027852>
4. <https://www.alliedmarketresearch.com/electrocardiograph-ECG-market>

How to cite this article:

Abhijit Ray.2023, New Screening Technique for Heart Failure. *International Journal of Current Advanced Research.* 12(08), pp. 2472-2473. DOI: <http://dx.doi.org/10.24327/ijcar.2023.2473.1542>
